

# EQPac

## Comprehensive Qualification

Model \_\_\_\_\_

Serial Number \_\_\_\_\_

Asset Number \_\_\_\_\_

Document ID \_\_\_\_\_

### Balance

## LabTec Instrument EQPac

## Table of Contents

<b>1. Introduction</b>	4.7 OQ Completion
1.1 Scope	<b>5. Performance Qualification (PQ)</b>
1.2 Contents	5.1 Routine tests by Balance User
1.3 Authorized Performer and Reviewer	5.2 Calibration by Service Provider
1.4 Equipment Qualification SOP	5.3 PQ Completion
1.5 Release of EQPac by Instrument Owner	<b>6. Completion and Release to Routine Operation</b>
1.6 Approval of (this section)	6.1 Completion of Installation
<b>2. Design Qualification (DQ)</b>	6.2 Release for Routine Operation
2.1 DQ Reference	<b>7. Maintenance Qualification (MQ)</b>
<b>3. Installation Qualification (IQ)</b>	7.1 Maintenance Provider
3.1 Reason for Qualification	7.2 Maintenance and/or Calibration Recommendation
3.2 Receiving and Unpacking	7.3 MQ Completion
3.3 Documents used for evaluation of completeness	
3.4 Instrument Identification and Completeness of Delivery	
3.5 Installation Site Verification	
3.6 Visual Verification of Equipment	
3.7 Start-up / Power-up Procedure	
3.8 System Location	
3.9 IQ Deviation Log	
3.10 IQ Completion	
<b>4. Operational Qualification (OQ)</b>	
4.1 Objective/Purpose	
4.2 Operational Readiness	
4.3 Functional Tests	
4.4 Initial Calibration Reports	
4.5 User Training	
4.6 OQ Deviation Log	